



College of American Pathologists
 325 Waukegan Road, Northfield, Illinois 60093-2750
 800-323-4040 • <http://www.cap.org>

Advancing Excellence

ORIGINAL EVALUATION

NB-C 2005 Neonatal Bilirubin Survey

INSTITUTION: Tanyalcin Tip Laboratuvari
 Izmir: TU 35220

ATTENTION: Orcan Tanyalcin MD

CAP NUMBER: 4317601-01 **Kit#** 1

KIT INFORMATION:	Kit ID: 18569955	Kit Mailed: 11/14/2005	Original Evaluation: 12/8/2005	Next Mailing Date: 4/10/2006
-------------------------	----------------------------	----------------------------------	--	--

COPIED TO: American Red Cross Natl Headqtrs CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
 <NONE>

 Reviewed By

 Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



College of American Pathologists
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • <http://www.cap.org>

Advancing Excellence

CAP Number: 4317601-01 Kit# 1
Institution: Tanyalcin Tip Laboratuvari
Attention: Orcan Tanyalcin MD
City / State: Izmir TU 35220

Kit ID: 18569955
Kit Mailed: 11/14/2005
Original Evaluation: 12/8/2005

EVALUATION
ORIGINAL

NB-C 2005 Neonatal Bilirubin Survey

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
Neonatal bilirubin mg/dL DIAZONIUM-DYPHYLL/NBIL VITROS DT60II CHEM SYS	NB-11				2					
	NB-12				1					
	NB-13				2					
	NB-14				2					
	NB-15				2					
All Method Principles	NB-11	10.4	10.53	0.64	1859	-0.2	8.4	12.7	Acceptable	
	NB-12	22.0	25.62	1.39	1843	-2.6	20.4	30.8	Acceptable	
	NB-13	5.3	5.39	0.35	1855	-0.2	4.3	6.5	Acceptable	
	NB-14	15.4	15.50	0.90	1852	-0.1	12.3	18.6	Acceptable	
	NB-15	9.5	10.53	0.64	1854	-1.6	8.4	12.7	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



College of American Pathologists
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • <http://www.cap.org>

Advancing Excellence

CAP Number: 4317601-01 Kit# 1
Institution: Tanyalcin Tip Laboratuvari
Attention: Orcan Tanyalcin MD
City / State: Izmir TU 35220

Kit ID: 18569955
Kit Mailed: 11/14/2005
Original Evaluation: 12/8/2005

EVALUATION
ORIGINAL

NB-C 2005 Neonatal Bilirubin Survey

CAP

CAP #: 4317601

Subspecialty: Routine Chemistry

Regulated Analyte	Proficiency Event 2005 1			Proficiency Event 2005 2			Proficiency Event 2005 3			Current Event Performance Interpretation	Cumulative CLIA '88 Performance Interpretation
	Test Event	Score	%	Test Event	Score	%	Test Event	Score	%		
Bilirubin, Total	NB-A	5/5	100	NB-B	5/5	100	NB-C	5/5	100	Satisfactory	Successful
Routine Chemistry		5/5	100		5/5	100		5/5	100	Satisfactory	Successful

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.